



## Mini Centrifuge User Manual

D1008



Please read carefully before use, and follow all operating and safety instructions!


### Specifications

Max. Speed [rpm]	7000
Max. RCF x g	2680
Rotor	1) 8x0.5/1.5/2.0ml Tubes 2) 16 x 0.2ml PCR tubes, 2x0.2ml PCR 8-tube Strip 3) 2 standard slides
Run Time	Continuous
Driving Motor	DC motor
Power	AC 100V~240V, 50Hz/60Hz
Noise Level	≤45 dB
Dimensions [DxWxH mm]	160 x 170 x 122
Weight [kg]	0.5

### Loading the Rotor

Place the rotor into position on the motor shaft. Press the rotor down to secure the rotor with shaft.

- Please use your both hands to load the rotor.



## **Operation**

- Open the lid and switch on the on/off switch is located back of the unit.
- Symmetrically load centrifuge tubes to rotor.
- Close the centrifuge lid to begin running.
- Switch off, the rotor will gradually stop.
- Open the lid and remove samples until the rotor extremely stopped.

## **Safety Precautions**

The detailed precautions described in the User Manual are carefully developed in an attempt to cover all the possible risks. Please read carefully before operating this centrifuge.

- Ensure that only trained staff works with the instrument.
- Never use the centrifuge in any manner not specified in these instructions.
- Never put hands in the rotor area unless the rotor is extremely stopped.
- If any abnormal condition occurs during operation, please stop it immediately and contact our service center.

## **Warranty**

This centrifuge and rotors are guaranteed for one year from the date of delivery provided that they have been operated and maintained properly. Please pay attention, do not use the rotor once it has been corrosion or fatigue damage.

## **Questions and Service**

In order to ensure to operate centrifuge safely and efficiently, it is necessary for regular maintenance. If centrifuge has problems, do not attempt to repair it by yourself. Contact our sales or service center for technical support in the following ways:

SCIOLOGEX, LLC  
1275 Cromwell Ave.  
Suite C6  
Rocky Hill, CT 06067 USA  
Tel: +1(860) 436-9221  
Fax: +1(860) 436-9745  
info@scilogex.com | www.scilogex.com